



### FAA Compliance for ResMed Devices

To whom it may concern,

The US Department of Transportation (DOT) Final Rule, "Nondiscrimination on the Basis of Disability in Air Travel" (73 FR 27614 which updates Title 14 CFR Part 382), effective May 13, 2009 provides important new requirements for the accommodation of passengers with respiratory assistive devices (Ventilators, Respirators and CPAP machines).

In line with these requirements, respiratory assistive devices may be used onboard an aircraft, without further testing by the carrier, provided they have been tested for Electromagnetic Compatibility (EMC) in accordance with the current version of RTCA/DO-160, Section 21, Category M.

<b>Product Code</b>	<b>Product Description</b>
33030	S8 Compact
33001, 33007, 33013, 33036, 33060	S8 Escape
33032	S8 Lightweight
33033	S8 AutoScore
33048, 33051, 33052, 33053, 33054, 33061	S8 Escape II
33055	S8 Lightweight II
33021	S8 Elite
33039, 33045, 33062, 33072, 33073, 33074, 33075	S8 Elite II
33044	S8 AutoScore II
33112	S8 AutoSet Vantage
33105	S8 AutoSet
33128, 33137	S8 Respond
33129, 33141, 33150	S8 AutoSet II
33136, 33143, 33144, 33145, 33146, 33147, 33148	S8 AutoSet Spirit II
33138	S8 AutoSet C
26101, 26121	VPAP Auto 25
26119, 26120	VPAP S
26110, 26122	VPAP ST
26104, 26105, 26125, 26126	S8 Auto 25
26102, 26103, 26106, 26107, 26108, 26109, 26123, 26127	VPAP IV
26112, 26113, 26114, 26115, 26116, 26117, 26118, 26124	VPAP IV ST
36003, 36013, 36023, 36103, 36113, 36123, 36143, 36203, 36223, 36243, 36263, 36303, 36343, 36363	S9 Elite Series
36005, 36015, 36025, 36105, 36115, 36125, 36145, 36205, 36225, 36245, 36265, 36305, 36325, 36345, 36365	S9 AutoSet Series
36065	S9 AutoSet for Her
36001, 36011, 36021, 36141, 36201, 36221, 36261, 36301, 36361	S9 Escape
36002, 36012, 36022, 36302, 36362	S9 Escape Auto
36004, 36014, 36024, 36114, 36124, 36144, 36204, 36224, 36284, 36304	S9 VPAP S (+ H5i, + Climate Control)
36006, 36016, 36026	S9 VPAP AUTO (+ H5i, + Climate Control)
36007, 36017, 36027, 36367	S9 VPAP ADAPT (+ H5i, + Climate

	Control)
36008, 36018, 36028, 36108, 36148, 36208, 36228, 36368	S9 VPAP ST (+ H5i, + Climate Control)
36100, 36110, 36120, 36140, 36200, 36220, 36360	S9 AUTOSET CS
36106, 36116, 36126, 36146, 36206, 36226, 36366	S9 AUTOSET 25
36118, 36128	S9 VPAP ST 22
24921, 24923, 24925	ResMed Power Station II (RPSII)

Some airlines may require notification greater than 48 hours before travel, and devices may need to be operated by battery. ResMed recommends that customers check with their airline.

ResMed has now successfully completed testing for the following respiratory assistive devices. **The devices listed below comply with RTCA/DO-160G, Section 21, Category M, and are considered FAA compliant.**



Simon Lewi  
Director – Regulatory Affairs, ResMed Ltd  
4 May, 2012